




Suivi de Mission

26/07/2010

Mission	Autoclave Ste	Id Enregistreur	0000001F31341 (DS1922T) /		
Début	28/09/2005 12:13:01 (TLS Toulouse)	intervalle	00:01:00	Min	70.0 °C
Fin	28/09/2005 15:03:01 (Atelier Auch)	Nb. Mesures	171	Max	100.0 °C

Notes				Stats	Min	Max	Moyenne	Dev St
					30.04 °C	120.98 °C	97.08 °C	22.00
Date Etal.		Etendue						



STERILISATION

Seuil Temp. Sté. : 121.1 °C
 Seuil Temp létale : 100 °C
 Seuil Temp. Past. : 70 °C
 Coeff. Z : 10
 VP / VS Cible : 3 mn
 Durée Mini : 0 mn
 VP : 3130835.18 mn
 VS : 24.14 mn
 Au dessus Seuil Sté : 0. mn
 Au dessus Seuil Past.: 146. mn
 Total observé : 171. mn

Stérilisation Réussie